

# EasyGo!™ Mouse FGFR3/CD333 One-Step ELISA Kit

[Catalog No] EK2F02EG


**[SIZE]** 48T/96T

**[ INTENDED USE ]** For the quantitative determination of Mouse fibroblast growth factor receptor 3 (FGFR3) concentrations in cell culture supernates, serum and plasma.

#### [INTRODUCTION]

FGFR3, also known as CD333, is a member of the fibroblast growth factor receptor (FGFR) family, with its amino acid sequence being highly conserved between members and among divergent species. FGFR family members differ from one another in their ligand affinities and tissue distribution. FGFRs are transmembrane catalytic receptors that have intracellular tyrosine kinase activity. Mutations in FGFR genes are the cause of several human developmental disorders characterized by skeletal abnormalities such as achondroplasia, and upregulation of FGFR expression may lead to cell transformation and cancer. FGFR3, a full-length representative protein would consist of an extracellular region, composed of three immunoglobulin-like domains, a single hydrophobic membrane-spanning segment and a cytoplasmic tyrosine kinase domain. The extracellular portion of FGFR3 interacts with fibroblast growth factors, setting in motion a cascade of downstream signals, ultimately influencing mitogenesis and differentiation. FGFR3 binds acidic and basic fibroblast growth hormone and plays a role in bone development and maintenance. CD333 is the receptor for acidic and basic fibroblast growth factors.

#### [PRINCIPLE OF THE ASSAY]





**( )** 

ELISA Plate Well Surface

This assay employs the quantitative sandwich enzyme immunoassay technique. A monoclonal antibody specific for Mouse FGFR3 has been immobilized onto microwells, and one pellet of the HRP-linked detect antibody specific for FGFR3 (light yellow) is pre-placed in the microwells, sealed by the adhesive film. Standard or samples are pipetted into the wells, then FGFR3 present is bound by the immobilized antibody and detect antibody in the incubation. After washing, substrate solution reacts with HRP and color develops in proportion to the amount of FGFR3 bound by the immobilized antibody. The color development is stopped and the intensity of the color is measured by microplate reader.

# [MATERIALS PROVIDED]

[MATERIALS FROVIDED]					
PART	PART#	EK2F02EG-	EK2F02EG-		
1740		48	96		
Coated Microplate	EK2F02EGP	48T	96T		
Standard	EK2F02EGS	1 vial	2 vials		
Assay Buffer (10×)	E0310	5 mL	5 mL		
TMB	E0230	6 mL	11 mL		
Stop Solution	E0300	11 mL	11 mL		
Washing Buffer (20×)	E0281	11 mL	11 mL		

Note: Components from reagent kits of different batch numbers must not be used interchangeably.

#### OTHER SUPPLIES REQUIRED

- 1) Microplate reader capable of measuring absorbance at 450 nm, with correction wavelength set at 570 nm or 630 nm.
- 2) Pipettes and pipette tips.
- 3) 50  $\,\mu\,L$  to 300  $\,\mu\,L$  adjustable multichannel micropipette with disposable tips.
- 4) Multichannel micropipette reservoir.
- 5) Beakers, flasks, cylinders necessary for preparation of reagents.
- 6) Deionized or distilled water.
- 7) Polypropylene test tubes for dilution.

#### [STORAGE]

Store at 2-8°C; refer to the kit label for expiration date.

For opened kits:

Pre-coated microplate: Can be stored at 2-8°C for approximately 1 month (return unused strips to the aluminum foil bag and reseal). Standard: Can be stored at -20°C for approximately 1 month (discard after single-use reconstitution).

Other components: Can be stored at 2-8°C for approximately 1 month.

## [SAMPLE COLLECTION AND STORAGE]

- 1) **Cell Culture Supernates** Remove particulates by centrifugation at  $300 \times g$  for 10 minutes and assay immediately or aliquot and store samples at  $\leq -20$ °C.
- 2) **Serum** Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 10 minutes at 1,000  $\times$  g. Remove serum and assay freshly prepared samples immediately or aliquot and store samples at  $\leq$  -20°C for later use. Avoid repeated freeze-thaw cycles.
- 3) **Plasma** Collect plasma using EDTA, citrate or heparin as anticoagulant. Centrifuge at 1,000 × g within 30 minutes of collection. Assay freshly prepared samples immediately or aliquot and store samples at ≤ -20°C for later use. Avoid repeated freeze-thaw cycles.
- 4) Other biological samples might be suitable for use in the assay. Serum and plasma were tested with this assay. Dilution with Assay Buffer (1x) may be needed.

Note: Samples containing a visible precipitate must be clarified prior to use in the assay. Do not use grossly hemolyzed or lipemic specimens

If samples are to be run within 24 hours, they may be stored at 2 to 8°C. For longer storage, aliquot samples and store frozen at -20°C. Avoid repeated freeze-thaw cycles.

## [REAGENT PREPARATION]

Bring all reagents and samples to room temperature before use. If crystals form in the Buffer Concentrates, warm and gently stir them until completely dissolved.

#### Washing Buffer (1x)

Pour entire contents (50 mL) of the **Washing Buffer (20×)** into a clean 1,000 mL graduated cylinder. Bring to final volume of 1,000 mL with pure or deionized water.

Mix gently to avoid foaming.

Transfer to a clean wash bottle and store at 2 to 25°C. Washing Buffer  $(1\times)$  is stable for 30 days.

## Assay Buffer (1x)

Pour the entire contents (5 mL) of the Assay Buffer (10×) into a clean 100 mL graduated cylinder. Bring to final volume of 50 mL with distilled water. Mix gently to avoid foaming.


Store at 2 to 8°C. Assay Buffer (1×) is stable for 30 days.

**Sample Dilution:** If your samples have high **FGFR3** content, dilute serum/plasma samples with Assay Buffer  $(1 \times)$ . For cell culture supernates, dilute with cell culture medium.


Mouse FGFR3 Standard: Reconstitute Mouse FGFR3 Standard by addition of distilled water. Reconstitution volume is stated on the label of the standard vial. Swirl or mix gently to insure complete and homogeneous solubilization (concentration of reconstituted standard = 20 ng/mL).

Allow the standard to reconstitute for 10 - 30 minutes. Mix well prior to making dilutions. Use polypropylene tubes.

- For serum/plasma samples, mixing concentrated Mouse FGFR3 standard (230 µL) with 230 µL of Assay Buffer (1x) creates the high standard (10 ng/mL). Pipette 230 µL of Assay Buffer (1x) into each tube. Use the high standard to produce a 1:1 dilution series (scheme below). Mix each tube thoroughly before the next transfer. Assay Buffer (1x) serves as the zero standard (0 pg/mL).
- For cell culture supernates, mixing concentrated Mouse FGFR3 standard (230 µL) with 230 µL of cell culture medium creates the high standard (10 ng/mL). Pipette 230 µL of cell culture medium into each tube. Use the high standard to produce a 1:1 dilution series. Mix each tube thoroughly before the next transfer. Cell culture medium serves as the zero standard (0 pg/mL).



## [ASSAY PROCEDURE]



Color developing and reading

Bring all reagents and samples to room temperature before use.

- 1) Prepare all reagents including microplate, samples, standards and working solution as described in the previous sections.
- 2) Remove excess microplate strips and return them to the foil pouch containing the desiccant pack, and reseal for further use. In any case, avoid touching the inner surface of the microwells and gently tap the plate to ensure that the pellets on the bottom of the microwells. Do not discard the pellets.
- 3) Adding Standard: Add 100  $\mu$ L of 2-fold diluted Standard to Standard well. Add 100  $\mu$ L of Assay Buffer (1×) to Blank well. The standards/samples can be added directly from the middle opening of the plate adhesive film.

- 4) Adding Samples: Serum/Plasma Add 80  $\mu$ L of Assay Buffer (1×) and 20  $\mu$ L sample to the sample well.Cell culture supernates Add 100  $\mu$ L cell culture supernates to the sample well.
- 5) **Incubation**: Incubate at 37°C for **1** hour, or at room temperature (25°C  $\pm$ 3°C) for **2** hours, on a microplate shaker set at 300-500 rpm.
- 6) **Washing**: Removing the plate adhesive film. Aspirate each well and wash by filling each well with 300µL of Washing Buffer (1×), repeating the process 3 times for a total four washes with 60 seconds interval. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
- 7) Adding Substrate Solution: Add 100 µL of Substrate Solution to each well. Incubate for 10±5 minutes at room temperature (25±3°C). Protect from light.
- 8) **Stopping**: Add 100  $\mu$ L of Stop Solution to each well. The color will turn yellow. If the color in the well is green or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
- 9) **Reading**: Measure the optical density value within 30 minutes by microplate reader set to 450 nm. If wavelength correction is available, set to 570 nm or 630 nm. If wavelength correction is not available, subtract readings at 570 nm or 630 nm from the readings at 450 nm. This subtraction will correct for optical imperfections in the plate. Reading directly at 450 nm without correction may generate higher concentration than true value.


#### [TYPICAL DATA]

A standard curve must be run within each assay. The following standard curve is provided for demonstration only.

Note: The finally concentration of top standard is 10 ng/mL.

If Serum/Plasma samples have been diluted following the instruction, the final dilution factor is 5. If sample have been diluted by other means, the concentration read from the standard curve must be multiplied by the appropriate dilution factor.

ng/mL	O.D.		Average	Corrected
0.00	0.006	0.009	800.0	
0.16	0.019	0.021	0.020	0.013
0.31	0.041	0.037	0.039	0.032
0.63	0.088	0.082	0.085	0.078
1.25	0.218	0.240	0.229	0.222
2.50	0.625	0.615	0.620	0.613
5.00	1.745	1.603	1.674	1.667
10.00	2.861	3.075	2.968	2.961



MULTI SCIENCES BIOTECH CO., LTD. www.multisciences.net E-mail: service@multisciences.net